

Altor Safety
711 Executive Boulevard
Valley Cottage, NY 10989
USA

The following sample(s) was/were submitted and identified by/on behalf of the client as: **Flat-fold disposable face masks**

Sample Received Date: **June 18, 2020**

Testing Period **June 18, 2020 – July 23, 2020**

Test Method & Results : Please refer to next page(s).

Test(s) Requested :

Test(s) Requested
ASTM F2100-19 – Standard Specification for Performance of Materials Used in Medical Face Masks
1. ASTM F2299/F2299M-03 (Reapproved 2017) – Test Method for Determining the Initial Efficiency of Materials used in Medical Face Masks to Penetration by Particulates Using Latex Spheres
2. EN14683:2019 Annex C Method for determination of breathability (differential pressure)
3. ASTM F1862/F1862M-17 – Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity)
4. 16 CFR Part 1610 – Standard for the Flammability of Clothing Textiles
5. ASTM F2101-19 – Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of <i>Staphylococcus aureus</i>

Signed for and on behalf of SGS North America, Inc.

Prepared By:




Daniel Miller
Air Labs Manager

Christina Crimi
Business Manager - Chemistry

This document is issued by the Company with the expectation that the document will be used internally only and not be used for any marketing or any other public use and is subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful, and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 7 days only.

As per ASTM F2100-19, Table 1 section 6.1


Medical Face Mask Material Requirements by Performance Level			
Test Parameter	Level 1 Barrier	Level 2 Barrier	Level 3 Barrier
Bacterial Filtration Efficiency (BFE)*	≥ 95%	≥ 98%	≥ 98%
Differential pressure	< 5.0 mmH ₂ O/cm ²	< 6.0 mmH ₂ O/cm ²	< 6.0 mmH ₂ O/cm ²
Sub-micron particulate filtration efficiency	≥ 95%	≥ 98%	≥ 98%
Resistance to penetration by synthetic blood	80 mmHg	120 mmHg	160 mmHg
Flame Spread	Class 1	Class 1	Class 1

Result Summary:

This table is a summary of results for all testing requested by the client. It is not to be used to state compliance and is only provided for ease of data consolidation. This report is not to be altered or reproduced unless in its entirety. The results are not valid unless accompanied by the data set(s) in the following Appendices.

Test Performed	Result Summary	Report	Appendix
Bacterial Filtration Efficiency (BFE)*	≥ 95%	Intertek Report No. 104369636CRT-012	1
Differential pressure	< 5.0 mmH ₂ O/cm ²	SGS IBR 22026C	2
Sub-micron particulate filtration efficiency	≥ 95%	SGS IBR 22026D	3
Resistance to penetration by synthetic blood	80 mmHg	SGS-IPS 01101-20	4
Flame Spread	Class 1	SGS-IPS 01101-20	4

*Testing is performed at an external laboratory.



SGS authenticates the photo(s) in the original report only

*** End of Report ***

This document is issued by the Company with the expectation that the document will be used internally only and not be used for any marketing or any other public use and is subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful, and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 7 days only.

APPENDIX 1



SGS IBR LABORATORIES TEST REPORT

SCOPE OF WORK

Performance Testing of Medical Face Masks to
ASTM F2100 *Standard Specification for Performance of
Materials Used in Medical Face Masks*, 2019 Edition

REPORT NUMBER

104369636CRT-012

ISSUE DATE

July 23, 2020

PAGES

4

DOCUMENT CONTROL NUMBER

GFT-OP-10i (28-Nov-2018)

© 2020 INTERTEK



This document is issued by the Company with the expectation that the document will be used internally only and not be used for any marketing or any other public use and is subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful, and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 7 days only.



Total Quality. Assured.

TEST REPORT

Issued July 23, 2020

 Intertek Report No. 104369636CRT-012
 Intertek Project No. G104369636

 3933 US Route 11
 Cortland, New York 13045

 Telephone: 1-607-753-6711
 Facsimile: 1-607-756-9891
 www.intertek.com

CLIENT

 SGS IBR Laboratories
 11599 Morrissey Rd
 Grass Lake, MI 49240
 USA

TEST STANDARD

 ASTM F2100 *Standard Specification for Performance of Materials Used in Medical Face Masks*, 2019 Edition

AUTHORIZATION

Quote Number: Qu-01081174-1

SAMPLE IDENTIFIED BY THE CLIENT AS

 Product Type: Face Mask
 Reference Number: JN 22026

SAMPLE INFORMATION

 Date(s) Samples Received: June 29, 2020
 Condition of Samples: Production Run
 Date(s) of Testing: July 21, 2020

TEST INFORMATION

ASTM F2101 <i>Bacterial Filtration Efficiency</i>	Test data attached
EN 14683:2019 Annex C <i>Differential Pressure</i>	Not tested under this project
ASTM F2299 <i>Sub-Micron Particulate Filtration</i>	Not tested under this project
ASTM F1862 <i>Resistance to Penetration by Synthetic Blood</i>	Not tested under this project
16 CFR 1610 <i>Flammability</i>	Not tested under this project

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.



Total Quality. Assured.

TEST REPORT

Issued 23-July-2020

Intertek Report No. 104369636CRT-012

Intertek Project No. G104369636

 3933 US Route 11
Cortland, New York 13045

 Telephone: 1-607-753-6711
Facsimile: 1-607-756-9891
www.intertek.com


**SECTION 1
CONCLUSION**

This test report represents the testing covered by proposal number Qu-01081174-1.

The observations and test results in this report are relevant only to the sample tested. Intertek makes no representations or warranties, express or implied, regarding units that were not tested including, but not limited to, units that may be part of the same lot.

If there are any questions regarding the results contained in this report, or any other services offered by Intertek, please do not hesitate to contact the undersigned.

Please note this Test Report does not represent authorization for the use of any Intertek certification marks.

Project Owner: Benjamin Hanna
Title: Project Engineer
Signature: 
Date: July 23, 2020

Project Reviewer: Jason Allen
Title: Technical Advisor
Signature: 
Date: July 23, 2020

REPORT REVISIONS

Date / Project #	Project Handler/ Reviewer	Description of Change
		None



Total Quality. Assured.

 SGS IBR Laboratories
 Intertek Report No: 104369636CART-012

TEST REPORT
SECTION 2
ASTM F2100-19 TEST DATA
BACTERIAL FILTRATION EFFICIENCY (BFE), ASTM F2101-19

Specimens conditioned for 4-hours at 20.4-22.1°C and 83-86%RH

Test Set-up Information	
Area of Test Specimen (cm ²)	48.3
Specimen Side Facing Challenge	Inside of Mask
Flow Rate (LPM)	28.3
Averaged + Control Plate Count	2942
Mean Particle Size (µm)	2.88, 2.93

SGS IBR JN 22026					
Plate Count	Mask Specimen				
Stage	1	2	3	4	5
Stage 1	1	1	0	0	0
Stage 2	0	0	1	0	0
Stage 3	1	0	0	2	0
Stage 4	7	5	3	4	6
Stage 5	16	5	5	12	11
Stage 6	0	3	1	0	0
Plate Count Total	25	14	10	18	17
% BFE	99.15	99.52	99.66	99.39	99.42


TEST EQUIPMENT INFORMATION

Description	Control Number	Calibration Date	Calibration Due
Conditioning Chamber	308-H252	2/26/2020	2/26/2021
Timer	308-H358	1/13/2020	1/13/2021
Pipette	308-H294	2/26/2020	2/26/2021
Analytical Balance	308-S268	12/2/2019	12/2/2020

Date of Testing	7/21/2020
-----------------	-----------

APPENDIX 2



IBR LABORATORIES

TEST REPORT

Test Method: EN14683:2019 Annex C Method for determination of breathability (differential pressure)

IBR JN: 22026C

Performed for: Altor Safety
Location: Valley Cottage, NY
Contact: Jared Soott

Date: 06 July 2020

Description of Samples: Flat-fold disposable face masks

Source: Seth MacGregor - Valley Cottage, NY
Date Samples Received: 18 June 2020

Fluid: Air
Flow Rate: 8 lpm
Effective Area: 4.9 cm²



Filter ID	Measurement Area	Differential Pressure (mmH ₂ O)	Mean Differential Pressure (mmH ₂ O)	Mean Differential Pressure / Area (mmH ₂ O/cm ²)	
22026-11	1	13.0	15.3	3.1	Temp: 23.6 °C RH: 45.5 % BP: 736 mmHg
	2	16.3			
	3	15.2			
	4	14.3			
	5	17.5			
22026-12	1	15.9	14.4	2.9	Temp: 23.6 °C RH: 46.8 % BP: 736 mmHg
	2	13.3			
	3	14.9			
	4	13.7			
	5	14.4			
22026-13	1	14.4	16.3	3.3	Temp: 23.6 °C RH: 45.5 % BP: 736 mmHg
	2	16.3			
	3	18.5			
	4	16.9			
	5	15.2			
22026-14	1	15.9	15.9	3.2	Temp: 23.5 °C RH: 45.4 % BP: 736 mmHg
	2	16.8			
	3	16.9			
	4	15.5			
	5	14.5			
22026-15	1	14.6	13.6	2.8	Temp: 23.6 °C RH: 45.5 % BP: 736 mmHg
	2	13.8			
	3	14.2			
	4	12.5			
	5	13.1			

Notice: These data relate only to the samples tested. This report may be copied only in its entirety.
Performed By: DN Data Location: DN262

Description	IBR ID	Manufacturer	Model No.	Serial No.	Range of Use	Cal Due
Flow Meter	AF-112	Alicat Scientific	M20SLPM-D/5M	99928	0.1-20 SLPM	8/30/2020
Differential Pressure	MAN-52	Dwyer	DHII-007	DC: A38AA	0.1-10.0 inH ₂ O	1/31/2021
Temp / Humidity	RH-206	Vaisala	HMT330	L5220038	12-75%RH/16-27C	1/9/2021
Barometric Pressure	MAN-51	Testo	511	39111389/505	300-1200 hPa	8/29/2020

Reviewed By: _____

Daniel R. Miller, Air Labs Manager

Revision	Editorial / Technical	Description	Approved By	Release Date
		Initial Release	DRM	7/7/2020

APPENDIX 3



TEST REPORT

Test Method: ASTM F2299/F2299M-03 (reapproved 2017) Determining the Initial Efficiency of Materials Used in Medical Face Masks to Penetration by Particulates Using Latex Spheres

Testing parameters per ASTM F2100-19 Standard Specification for Performance of Materials Used in Medical Face Masks

IBR JN: 22026D

Date: 06 July 2020

Description of Samples: Flat-fold disposable face masks

Test Area: 45.22 cm²

Source: Seth MacGregor - Valley Cottage, NY

Date Samples Received: 18 June 2020

Fluid: Air

Flow Rate: 28.3 lpm

Face Velocity: 10.4 cm/s

Challenge: 0.1µm (±15% CV) Latex Microspheres (Neutralized)

Performed for: Altor Safety

Location: Valley Cottage, NY

Contact: Jared Scott



Filter ID	Differential Pressure (mmH ₂ O)	Port	Particles / 2 ft3	Temp:	RH:	BP:
22026-1	6.6	Upstream	6242875	21.3 °C	49.9 %	735 mmHg
		Downstream	274922			
		Efficiency (%)	95.60			
22026-2	6.9	Upstream	7394750	21.4 °C	49.6 %	735 mmHg
		Downstream	270156			
		Efficiency (%)	96.35			
22026-3	6.4	Upstream	6869425	21.2 °C	49.7 %	735 mmHg
		Downstream	286204			
		Efficiency (%)	95.83			
22026-4	7.1	Upstream	6813800	21.2 °C	49.4 %	735 mmHg
		Downstream	296737			
		Efficiency (%)	95.65			
22026-5	7.1	Upstream	6896900	21.4 °C	49.0 %	735 mmHg
		Downstream	265120			
		Efficiency (%)	96.16			

Notice: These data relate only to the samples tested. This report may be copied only in its entirety.

Performed By: DN

Data Location: DN262

Manufacturer	Model Number	Serial Number	IBR ID	Range of Use	Cal Due
Alicat Scientific	M-50SLPM-D/5M	99929	AF-113	5-45 SLPM	9/3/2020
Dwyer	DHII-007	Date Code: A31X	MAN-31	0.1-10.0 inH ₂ O	2/17/2021
Vaisala	HMT330	L5220038	RH-206	12-75%RH/16-27C	1/9/2021
Vaisala	PTU300	R3240750	RH-209	500-1100 hPa	8/9/2020
PMS	Lasair III 110	116514	N/A	0.1-5.0 µm	12/17/2020
PMS	Lasair III 110	102709	N/A	0.1-5.0 µm	9/1/2020

Reviewed By: _____

Daniel R. Miller, Air Labs Manager

Revision	Editorial / Technical	Description	Approved By	Release Date
		Initial release	DRM	7/7/2020

SGS IBR Laboratories 11599 Morrissey Rd Grass Lake MI 49240 USA

Voice: +1 517 522 8453

page 1 of 1

This document is issued by the Company with the expectation that the document will be used internally only and not be used for any marketing or any other public use and is subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful, and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 7 days only.

APPENDIX 4

Report to: Jared Scott
Altor Safety
711 Executive Boulevard
Valley Cottage, NY 10989

Sample Identification: Project JN 22026 (One Face Mask Sample)

Date Received: July 1, 2020


Test(s) Requested: Flammability, Resistance to Blood Penetration - ASTM F 1862


PO Number: SGS IBR (JN 22026)

Analysis of One Face Mask Sample

SGS IPS Testing performed the testing listed above on one face mask sample provided by Altor Safety. The results are summarized in Tables 1 and 2 on the following pages.

If you have any questions, please contact us.

Authorized by 
MaiPa Xiong
Laboratory Manager

Signed 
Andrew Smits
Senior Lab Technician
Physical Testing
920-749-3040

Report to Altor Safety
 SGS-IPS 01101-20

Project No. JN 22026

 July 17, 2020
 Page 2 of 6

Table 1. Physical Properties

 JN 22026 Face
 Masks

Flammability


AT-1659

Orientation	Nose to chin
Test Side	Outside
Burn Time (sec)	
1	IBE
2	IBE
3	IBE
4	IBE
5	IBE
Average	IBE

Classifications Class 1



IBE indicates the specimen Ignited But Extinguished.
 DNI indicates the specimen Did Not Ignite.
 Average calculated based on combusted samples only.

Report to Altor Safety
 SGS-IPS 01101-20

Project No. JN 22026

 July 17, 2020
 Page 3 of 6

Table 2. Resistance to Blood Penetration - ASTM F 1862

	JN 22026 Face Masks (160mmHg)	JN 22026 Face Masks (120mmHg)	JN 22026 Face Masks (80mmHg)
	  AT-1659		
Resistance to Blood Penetration (Pass/Fail)			
1	Fail	Pass	Pass
2	Fail	Pass	Pass
3	Fail	Pass	Pass
4	Fail	Pass	Pass
5	NA	Fail	Pass
6	NA	Pass	Pass
7	NA	Pass	Pass
8	NA	Pass	Fail
9	NA	Pass	Pass
10	NA	Pass	Pass
11	NA	Pass	Pass
12	NA	Fail	Pass
13	NA	Fail	Fail
14	NA	Pass	Pass
15	NA	Fail	Pass
16	NA	NA	Pass
17	NA	NA	Pass
18	NA	NA	Pass
19	NA	NA	Pass
20	NA	NA	Pass
21	NA	NA	Pass
22	NA	NA	Pass
23	NA	NA	Pass
24	NA	NA	Pass
25	NA	NA	Pass
26	NA	NA	Pass
27	NA	NA	Pass
28	NA	NA	Pass
29	NA	NA	Pass
30	NA	NA	Pass
31	NA	NA	Pass
32	NA	NA	Pass
Resistance to Blood Penetration Summary	0 Pass/4 Fail	11 Pass/4 Fail	30 Pass/2 Fail
Test Parameters			
Test Pressure (mmHg)	160	120	80
Lot Number	47201	47201	47201

Report to Altor Safety
 SGS-IPS 01101-20

Project No. JN 22026

 July 17, 2020
 Page 4 of 6

Method(s) and Notes:

All valid results are included in the statistical analyses.
 Revisions of SGS-IPS methods when used are current at the time of testing.
 Samples tested and conditioned in TAPPI standard conditions unless requested otherwise by customer.
 Samples were not preconditioned.

16 CFR Ch. II Part 1610 (1-1-18 Edition) - Standard for the Flammability of Clothing Textiles

Testing conducted as a Plain Surface Textile Fabric.
 Samples were not refurbished.

Preliminary tests to determine most rapidly burning direction were not conducted.
 Samples were cooled in a desiccator after oven drying for a minimum of 15 minutes.
 If samples could not be dried in an oven, samples were placed in a desiccator for a minimum of 1 hour.

ASTM F 1862/F 1862M-17 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity)

A 2 mL check was performed at the beginning, middle and end of each sample.
 Synthetic blood is purchased from Johnson, Moen & Co. Surface Tension is not independently verified after receipt and unused synthetic blood is stored in original plastic bottles.

As requested, the sample was initially tested using the highest pressure of 160mmHg. At 160mmHg, the sample failed to meet the Acceptable Quality Limit (AQL) specified in ASTM F 1862. Testing then proceeded to 120mmHg and failed to meet the AQL. Testing then proceeded to 80mmHg which met the AQL.

 Analyzed by: EJW, EB, RG

 Quality review by: TFH

 Date(s) of testing: July 7-17, 2020
Room Conditions

	Relative Humidity (%)	Temperature (°F)
Conditioning Environment	50.9	73.6
Maximum during testing	51.8	73.9
Minimum during testing	50.9	73.6

Note: See the method(s) cited above for available estimates of measurement uncertainty. Unless otherwise noted, sampling was performed by customer.

Report to Altor Safety
SGS-IPS 01101-20

Project No. JN 22026

July 17, 2020
Page 5 of 6

This document is issued by the Company with the expectation that the document will be used internally only and not be used for any marketing or any other public use and is subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



This document is issued by the Company with the expectation that the document will be used internally only and not be used for any marketing or any other public use and is subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/terms_and_conditions.htm and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful, and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 7 days only.

Report to Altor Safety
SGS-IPS 01101-20

Project No. JN 22026

July 17, 2020
Page 6 of 6**Parameters and Instruments Appendix**
Resistance to Blood Penetration - ASTM F 1862Chamber Humidity (%RH)
81.97Chamber Temperature (°F)
70.22